



CATALOGUE

ALINITY LINE





Company profile

MISSION STATEMENT

«To have a perfect control of Development, Manufacturing process & Selling of In Vitro Diagnostic liquid reagents with a strong focus on our Customer vision»

Your R&D Partner

DiAgam manufactures **reagents for immunoturbidimetry**.

Our company specializes in high technology R&D, based on immunology technologies. DiAgam is the creator of liquid laboratory reagents containing **gold particles**.

We are currently taking part in a project to develop **new nanotechnologies** based on the use of **colloidal particles** in the field of medical diagnosis. This activity has also succeeded in completing our products range with Cystatin C, Ferritin & Rheumatoid factor.

Manufacturer of ready-to-use reagents

All our products are developed and manufactured within the company.

Our reagents, controls and calibrators are liquid, stable and ready to use on **all opened biochemistry analyzers even with barcodes** (Abbott - Alinity – Architect , Beckman Coulter – AU series, IDS-iSYS Multi-Discipline Automated System, Menarini – Zenit RA, Mindray BS series, Ortho - VITROS® 5,1 FS Chemistry, VITROS® 4600 Chemistry, VITROS® 5600 Integrated Systems, Roche - Modular®-COBAS C, Siemens - Advia® 1800 - Advia® 2400 - Atellica®, Thermo-Konelab®, etc...)

DiAgam is a member of the National Association of **European Diagnostic Manufacturers Association (EDMA)**. All reagents are labelled and DiAgam is certified **ISO 9001:2015** and **ISO 13485:2016**.

All calibrators and controls are traceable to the **International Standards** (WHO, IFCC, NIBSC, etc.)

Worldwide reliability

DiAgam promotes its products in B to B and B to C businesses through its direct affiliates in Belgium, France and Iberia. We also export in ASEAN, EMAE, Americas.

Thanks to our **manufacturing flexibility**, we can propose all our parameters in three different packagings: **Kits** especially dedicated to particular analyzers, **Vials** filled and **Bulk**, available under **DiAgam** label or in **OEM**.


We respect our philosophy by offering our customer, a **high quality of our products** with a most **competitive price** along with a continuous quest to **satisfy their needs**.

ALBUMIN
ALPHA 1 - ACID GLYCOPROTEIN
ALPHA 1 - ANTITRYPSIN
ALPHA 1 - MICROGLOBULIN
ALPHA 2 - MACROGLOBULIN
ANTITHROMBIN III
APOLIPOPROTEIN A1
APOLIPOPROTEIN B
ASO

BETA 2 - MICROGLOBULIN
CALPROTECTIN *(new)*
CERULOPLASMIN
C3
C4
CRP
CYSTATIN C
FERRITIN
FIBRINOGEN
HAPTOGLOBIN

Ig A
Ig E *(new)*
Ig G
Ig M
LIPOPROTEIN A
MICROALBUMIN
PREALBUMIN
RETINOL BINDING PROTEIN
RHEUMATOID FACTOR
TRANSFERRIN





CATALOGUE

Calprotectin **New!**

<i>REAGENTS</i>	<i>Reference</i>	<i>Content</i>	<i>Tests*</i>	<i>Shelf Life**</i>	<i>Sample***</i>	<i>Analytical Range****</i>
-----------------	------------------	----------------	---------------	---------------------	------------------	-----------------------------

Please contact us for more information

CALPROTECTIN TEST

SARS-CoV-2 SEROLOGICAL ASSAY **New!**

<i>REAGENTS</i>	<i>Reference</i>	<i>Content</i>	<i>Tests*</i>	<i>Shelf Life*</i>	<i>Sample**</i>	<i>Analytical Range</i>
-----------------	------------------	----------------	---------------	--------------------	-----------------	-------------------------

Please contact us for more information

SARS-CoV-2 SEROLOGICAL ASSAY based on PETIA Colloidal Gold

CATALOGUE

ALBUMIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ALTUR-B00/ALI	1 x 20 ml R1 + 1 x 7,5 ml R2		100	36 months	serum or plasma	1,096 to 120,44 (g/L)
	ALTUR-H00/ALI	1 x 74 ml R1 + 1 x 19,5 ml R2		343			
	ALTUR-L00/ALI	10 x 74 ml R1 + 10 x 19,5 ml R2		3435			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALBUMIN CSF

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	MATUR-B00/ALI	1 x 22,5 ml R1 + 1 x 7 ml R2		100	36 months	cerebrospinal fluid	10,86 - 494,33 mg/L
	MATUR-H00/ALI	1 x 70 ml R1 + 1 x 17,6 ml R2		342			
	MATUR-L00/ALI	5 x 70 ml R1 + 5 x 17,6 ml R2		1712			

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	MACOZ-002	1x2 ml	18 months
H	MACOX-002	1x2 ml	18 months
M	MACON-002	1x2 ml	18 months
L	MACOS-002	1x2 ml	18 months

ALPHA-1 ACID-GLYCOPROTEIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	GPTUR-B00/ALI	1 x 30,5 ml R1 + 1 x 8,6 ml R2		100	36 months	serum	0,143 - 2,42 g/L
	GPTUR-H00/ALI	1 x 70 ml R1 + 1 x 17,2 ml R2		243			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

ALPHA-1 ANTITRYPSIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ATTUR-B00/ALI	1 x 31 ml R1 + 1 x 8,6 ml R2		100	36 months	serum or plasma	0,007 to 3,5 (g/L)
	ATTUR-H00/ALI	1 x 70 ml R1 + 1 x 17,1 ml R2		242			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-1 MICROGLOBULIN (URINE)

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	A1DGX-B00/ALI	1 x 14,5 ml R1 + 1 x 14 ml R1		100	6 months	urine	1,29 to 166,6 (mg/L)
	A1DGX-H00/ALI	1 x 46 ml R1 + 1 x 45,5 ml R2		376			

CALIBRATORS

KIT 5 LEV	A1REK-000	5x1 ml	6 months
-----------	-----------	--------	----------

CONTROLS

M	A1CON-002	1x2 ml	6 months
L	A1COS-002	1x2 ml	6 months

ALPHA-2 MACROGLOBULIN - ALPHA-2 MACROGLOBULIN URINARY

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	MGTUR-B00/ALI	1 x 30,5 ml R1 + 1 x 8,6 ml R2		137	36 months	serum or plasma or urine	0,026 to 4,96 (g/L) serum 3,36 - 199 (mg/L) urine
	MGTUR-H00/ALI	1 x 74 ml R1 + 1 x 18 ml R2		352			
	MGTUR-L00/ALI	5 x 74 ml R1 + 5 x 18 ml R2		1758			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months
URINARY LOW	MGCU1-002	1x2 ml	12 months
URINARY HIGH	MGCU2-002	1x2 ml	12 months

CATALOGUE

ANTI-STREPTOLYSIN O

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	ASLTX-B00/ALI	1 x 29 ml R1 + 1 x 6,1 ml R1	100	20 months	serum or plasma	19,99 to 864,85 (IU/ml)
	ASLTX-H00/ALI	1 x 70 ml R1 + 1x 11,7 ml R2	259			

CALIBRATORS

KIT 5 LEV	ASREK-000	5x1 ml	12 months
-----------	-----------	--------	-----------

CONTROLS

M	ASCON-002	1x2 ml	12 months
---	-----------	--------	-----------

ANTITHROMBIN III

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	A3TUR-B00/ALI	1 x 25,5 ml R1 + 1 x 7,5 ml R1	100	36 months	plasma	28,15 - 806,3 mg/L
	A3TUR-H00/ALI	1 x 70 ml R1 + 1x 17,4 ml R2	301			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

APOLIPOPROTEIN A1

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	AATUR-B00/ALI	1 x 31 ml R1 + 1 x 10,4 ml R1	101	36 months	serum or plasma	0,029 - 4 g/L
	AATUR-H00/ALI	1 x 70 ml R1 + 1x 21,4 ml R2	243			

CALIBRATORS

KIT 5 LEV	ABREK-000	5 x 0,5 ml	18 months
-----------	-----------	------------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	2-8 °C	18 months
M	ABCON-002	1x2 ml	2-8 °C	18 months

CATALOGUE

APOLIPOPROTEIN B

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	ABTUR-B00/ALI	1 x 21,6 ml R1 + 1 x 6,7 ml R1	98	36 months	serum or plasma	0,019 - 4 g/L
	ABTUR-H00/ALI	1 x 70 ml R1 + 1x 17,6 ml R2	360			

CALIBRATORS

KIT 5 LEV	ABREK-000	5 x 0,5 ml	18 months
-----------	-----------	------------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

BETA2-MICROGLOBULIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	B2DGX-B00/ALI	1 x 23 ml R1 + 1 x 11,8 ml R2	100	15 months	serum or plasma	0,132 - 95 mg/L
	B2DGX-H00/ALI	1 x 70 ml R1 + 1 x 33,5 ml R2	334			
	B2DGX-L00/ALI	5 x 70 ml R1 + 5 x 33,5 ml R2	1672			

CALIBRATORS

KIT 5 LEV	B2REK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	B2COX-001	1x1 ml	18 months
M	B2CON-003	1x3 ml	18 months
L	B2COS-003	1x3 ml	18 months

BETA2-MICROGLOBULIN URINARY

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	BUDGX-B00/ALI	1 x 18,2 ml R1 + 1 x 9,7 ml R2	100	Contact us	Urine	0,055 - 9,3 mg/L
	BUDGX-H00/ALI	1 x 74 ml R1 + 1 x 35,7 ml R2	468			

CALIBRATORS

KIT 5 LEV	BUREK-000	5x1 ml	Contact us
-----------	-----------	--------	------------

CONTROLS

M	B2CON-003	1x3 ml	Contact us
L	B2COS-003	1x3 ml	Contact us

CALPROTECTIN (in feces)

AVAILABLE IN 2021

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	CACOL-B00/ALI	AVAILABLE IN 2021	100	Contact us	Feces	Contact us
	CACOL-H00/ALI	AVAILABLE IN 2021	300			

CALIBRATORS

KIT 5 LEV	CAREK-000	6x1 ml	Contact us
-----------	-----------	--------	------------

CONTROLS

L/M/H	Contact us	contact us	Contact us
-------	------------	------------	------------

CATALOGUE

CERULOPLASMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	CETUR-B00/ALI	1 x 25,1 ml R1 + 1 x 7,5 ml R2	100	36 months	serum or plasma	9,48 - 1051,2 mg/L
	CETUR-H00/ALI	1 x 70 ml R1 + 1x 17,4 ml R2	301			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C3

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	C3TUR-B00/ALI	1 x 24 ml R1 + 1 x 7,2 ml R2	100	36 months	serum or plasma	0,015 - 4,31 g/L
	C3TUR-H00/ALI	1 x 70 ml R1 + 1x 17,4 ml R2	322			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C4

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	C4TUR-B00/ALI	1 x 25,1 ml R1 + 1 x 7,5 ml R2	100	36 months	serum or plasma	0,002 - 0,825 g/L
	C4TUR-H00/ALI	1 x 70 ml R1 + 1x 17,4 ml R2	301			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

CRP GOLD LATEX

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	CPDGX-B00/ALI	1 x 14,5 ml R1 + 1 x 14,1 ml R2	100	24 months	serum or plasma	0,031 - 300 mg/L
	CPDGX-H00/ALI	1 x 46 ml R1 + 1 x 45,5 ml R2	376			
	CPDGX-L00/ALI	5 x 46 ml R1 + 5 x 45,5 ml R2	1882			

CALIBRATORS

KIT 5 LEV	CPRLK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

CRP WIDE RANGE

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	CPTXL-B00/ALI	1 x 31 ml R1 + 1 x 7,1 ml R2	100	36 months	serum or plasma	0,607 - 425 mg/L
	CPTXL-H00/ALI	1 x 74 ml R1 + 1 x 14,1 ml R2	256			
	CPTXL-L00/ALI	5 x 74 ml R1 + 5 x 14,1 ml R2	1281			

CALIBRATORS

KIT 5 LEV	CPREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

CYSTATIN C

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	CYCOL-B00/ALI	1 x 23 ml R1 + 1 x 11,3 ml R2	100	6 months	serum or plasma	0,15 - 9,355 mg/L
	CYCOL-H00/ALI	1 x 70 ml R1 + 1 x 31,7 ml R2	334			
	CYCOL-L00/ALI	5 x 70 ml R1 + 5 x 31,7 ml R2	1672			

CALIBRATORS

KIT 5 LEV	CYREK-000	5x1 ml	12 months
-----------	-----------	--------	-----------

CONTROLS

H	CYCOX-002	1x2 ml	12 months
M	CYCON-002	1x2 ml	12 months
L	CYCOX-002	1x2 ml	12 months

CATALOGUE

FERRITIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	FNCOL-B00/ALI	1 x 25,1 ml R1 + 1 x 7,5 ml R2	100	12 months	serum	6,344 - 500 µg/L
	FNCOL-H00/ALI	1 x 70 ml R1 + 1 x 17,4 ml R2	301			

CALIBRATORS

KIT 5 LEV	FNRGK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

M	FNCON-002	1x2 ml	18 months
---	-----------	--------	-----------

FIBRINOGEN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	FITUR-B00/ALI	1 x 31 ml R1 + 1 x 8,6 ml R2	100	36 months	plasma (from sodium citrate)	0,05 - 10 g/L
	FITUR-H00/ALI	1 x 70 ml R1 + 1 x 17,2 ml R2	243			

CALIBRATORS

KIT	FIREM-001	1x1 ml	12 months
-----	-----------	--------	-----------

CONTROLS

M	FICON-002	1x2 ml	12 months
---	-----------	--------	-----------

HAPTOGLOBIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	HATUR-B00/ALI	1 x 30,5 ml R1 + 1 x 8,6 ml R2	100	36 months	serum or plasma	0,002 - 3,86 g/L
	HATUR-H00/ALI	5 x 70 ml R1 + 5 x 17,2 ml R2	1214			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS A

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	IATUR-B00/ALI	1 x 21,5 ml R1 + 1 x 7 ml R2	100	36 months	serum	0,02 - 8 g/L
	IATUR-H00/ALI	1 x 70 ml R1 + 1 x 18,5 ml R2	363			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months

L	MPCOS-002	1x2 ml	24 months
---	-----------	--------	-----------

CATALOGUE

IMMUNOGLOBULINS E

NEW

Contact us

IMMUNOGLOBULINS G

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IGTUR-B00/ALI	1 x 21,6 ml R1 + 1 x 10,1 ml R2		100	36 months	serum	0,125 to 35 (g/L)
	IGTUR-H00/ALI	1 x 70 ml R1 + 1 x 29,5 ml R2		358			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS G (CSF)

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IGLCR-B00/ALI	1 x 18,5 ml R1 + 1 x 5,9 ml R2		100	36 months	cerebrospinal fluid	3,375 to 249,78 (mg/L)
	IGLCR-H00/ALI	1 x 74 ml R1 + 1 x 18 ml R2		467			

CALIBRATORS

KIT 5 LEV	IGREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	IGCOX-002	1x2 ml	24 months
M	IGCON-002	1x2 ml	24 months
L	IGCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS M

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IMTUR-B00/ALI	1 x 21,6 ml R1 + 1 x 7 ml R2		100	36 months	serum	0,038 - 7 g/L
	IMTUR-H00/ALI	1 x 70 ml R1 + 1 x 18,5 ml R2		363			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

LIPOPROTEIN A

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
		<i>R1</i>	<i>R2</i>			
KIT	LPTUR2-B00/ALI	1 x 25,5 ml R1 + 1 x 5,9 ml R2	100	12 months	serum or plasma	0,0235 - 0,946 g/L
	LPTUR2-H00/ALI	1 x 70 ml R1 + 1 x 12,5 ml R2	300			

CALIBRATORS

KIT 5 LEV	LPREK-000	4 x 0,5 ml	10 months
-----------	-----------	------------	-----------

CONTROLS

M	LPCON-002	1 x 2 ml	12 months
---	-----------	----------	-----------

MICROALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
		<i>R1</i>	<i>R2</i>			
KIT	MATUR-B00/ALI	1 x 22,5 ml R1 + 1 x 7 ml R2	100	36 months	urine	1,860 - 400 mg/L
	MATUR-H00/ALI	1 x 70 ml R1 + 1 x 17,6 ml R2	342			
	MATUR-L00/ALI	5 x 70 ml R1 + 5 x 17,6 ml R2	1712			

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	MACOX-002	1x2 ml	18 months
M	MACON-002	1x2 ml	18 months
L	MACOS-002	1x2 ml	18 months

PREALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	PATUR-B00/ALI	1 x 30,5 ml R1 + 1 x 8,6 ml R2	100	36 months	serum or plasma	6,53 to 1000 (mg/L)
	PATUR-H00/ALI	1 x 74 ml R1 + 1 x 18 ml R2	257			
	PATUR-L00/ALI	5 x 74 ml R1 + 5 x 18 ml R2	1283			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

RETINOL BINDING PROTEIN (RBP)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	RBTUR-B00/ALI	1 x 20 ml R1 + 1 x 8,6 ml R2	100	24 months	serum or plasma	7,81 to 141 (mg/L)
	RBTUR-H00/ALI	1 x 70 ml R1 + 1 x 26,5 ml R2	398			
	RBTUR-L00/ALI	5 x 70 ml R1 + 5 x 26,5 ml R2	1992			

CALIBRATORS

KIT 5 LEV	RBREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	RBCOX-002	1x2 ml	24 months
M	RBCON-002	1x2 ml	24 months
L	RBCOS-002	1x2 ml	24 months

RHEUMATOID FACTOR

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	RFCOL-B00/ALI	1 x 25,5 ml R1 + 1 x 17,5 ml R2	100	18 months	serum or plasma	2,228 - 120 kIU/L
	RFCOL-H00/ALI	1 x 68 ml R1 + 1 x 46 ml R2	290			

CALIBRATORS

KIT 5 LEV	RFREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	RFCOX-002	1x2 ml	18 months
M	RFCON-002	1x2 ml	24 months
L	RFCOS-002	1x2 ml	18 months

TRANSFERRIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	TFTUR-B00/ALI	1 x 31 ml R1 + 1 x 10,8 ml R2	100	36 months	serum or plasma	0,076 - 8 g/L
	TFTUR-H00/ALI	1 x 70 ml R1 + 1 x 22,4 ml R2	243			
	TFTUR-L00/ALI	5 x 70 ml R1 + 5 x 22,4 ml R2	1213			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months



Your expert in specific proteins

-
- (*) Tests :** Amount of real tests. May vary from a analyzer to another one.
- (**) Shelf Life :** DiAgam provides to distributors a minimum of **12 months** .(except for cystatin C, lipoprotein (a) and Alpha1-Microglobulin). The time period written in this document shows the maximum shelf life time after manufacturing.
- (***) Sample :** DiAgam may provide some applications for other samples, such as CSF, Urine, Veterany samples etc.... Please contact your distributor or us directly at mail@diagam.com or dial **+32 68 55 14 82**.
- (****) Analytical range :** For general information purposes only. More details can be found in the package insert (instructions for use). Please contact your distributor or us directly at mail@diagam.com or dial **+32 68 55 14 82** in order to get the updated data.



Medical device for in vitro diagnosis.



DIAGAM SA, *Rue du Parc Industriel 40, 7822 Ghislenghien, BELGIUM*



The products fulfill the requirements of European directive 98/79 EC concerning the medical devices of in vitro diagnosis.

© DiAgam. 2021. All rights reserved. All product names, registered trademarks, company names in this document remain the property of their respective owners. For general information purposes only. More details can be found in the package insert (instructions for use) and labelling. Not for the US market except for research use only.