



CATALOGUE

ATELLICA LINE





Company profile

MISSION STATEMENT

«To have a perfect control of Development, Manufacturing process & Selling of In Vitro Diagnostic liquid reagents with a strong focus on our Customer vision»

Your R&D Partner

DiAgam manufactures **reagents for immunoturbidimetry**.

Our company specializes in high technology R&D, based on immunology technologies. DiAgam is the creator of liquid laboratory reagents containing **gold particles**.

We are currently taking part in a project to develop **new nanotechnologies** based on the use of **colloidal particles** in the field of medical diagnosis. This activity has also succeeded in completing our products range with Cystatin C, Ferritin & Rheumatoid factor.

Manufacturer of ready-to-use reagents

All our products are developed and manufactured within the company.

Our reagents, controls and calibrators are liquid, stable and ready to use on **all opened biochemistry analyzers even with barcodes** (Abbott - Alinity – Architect, Beckman Coulter – AU series, IDS-iSYS Multi-Discipline Automated System, Menarini – Zenit RA, Mindray BS series, Ortho - VITROS® 5,1 FS Chemistry, VITROS® 4600 Chemistry, VITROS® 5600 Integrated Systems, Roche - Modular®-COBAS C, Siemens - Advia® 1800 - Advia® 2400 - Atellica®, Thermo-Konelab®, etc...)

DiAgam is a member of the National Association of **European Diagnostic Manufacturers Association (EDMA)**. All reagents are labelled and DiAgam is certified **ISO 9001:2015** and **ISO 13485:2016**.

All calibrators and controls are traceable to the **International Standards** (WHO, IFCC, NIBSC, etc.)

Worldwide reliability

DiAgam promotes its products in B to B and B to C businesses through its direct affiliates in Belgium, France and Iberia. We also export in ASEAN, EMAE, Americas.

Thanks to our **manufacturing flexibility**, we can propose all our parameters in three different packagings: **Kits** especially dedicated to particular analyzers, **Vials** filled and **Bulk**, available under **DiAgam** label or in **OEM**.


We respect our philosophy by offering our customer, a **high quality of our products** with a most **competitive price** along with a continuous quest to **satisfy their needs**.

ALBUMIN
ALPHA 1 - ACID GLYCOPROTEIN
ALPHA 1 - ANTITRYPSIN
ALPHA 1 - MICROGLOBULIN
ALPHA 2 - MACROGLOBULIN
ANTITHROMBIN III
APOLIPOPROTEIN A1
APOLIPOPROTEIN B
ASO

BETA 2 - MICROGLOBULIN
CALPROTECTIN *(new)*
CERULOPLASMIN
C3
C4
CRP
CYSTATIN C
FERRITIN
FIBRINOGEN
HAPTOGLOBIN

Ig A
Ig E *(new)*
Ig G
Ig M
LIPOPROTEIN A
MICROALBUMIN
PREALBUMIN
RETINOL BINDING PROTEIN
RHEUMATOID FACTOR
TRANSFERRIN





CATALOGUE

New

Calprotectin

<i>REAGENTS</i>	<i>Reference</i>	<i>Content</i>	<i>Tests*</i>	<i>Shelf Life**</i>	<i>Sample***</i>	<i>Analytical Range****</i>
-----------------	------------------	----------------	---------------	---------------------	------------------	-----------------------------

Please contact us for more information

CALPROTECTIN TEST

New

SARS-CoV-2 SEROLOGICAL ASSAY

<i>REAGENTS</i>	<i>Reference</i>	<i>Content</i>	<i>Tests*</i>	<i>Shelf Life**</i>	<i>Sample***</i>	<i>Analytical Range****</i>
-----------------	------------------	----------------	---------------	---------------------	------------------	-----------------------------

Please contact us for more information

SARS-CoV-2 SEROLOGICAL ASSAY based on PETIA Colloidal Gold

CATALOGUE

ALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	ALTUR-B00/ATE	1 x 1 x 13,2 ml R1 + 1 x 1 x 5,2 ml R2	101	36 months	serum or plasma	0.815 to 124.94 (g/l)
	ALTUR-H00/ATE	1 x 2 x 24,5 ml R1 + 1 x 2 x 8,7 ml R2	395			
	ALTUR-L00/ATE	5 x 2 x 24,5 ml R1 + 5 x 2 x 8,7 ml R2	1973			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
------------------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALBUMIN CSF

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
KIT	MATUR-B00/ATE	1 x 1 x 13,1 ml R1 + 1 x 1 x 4,6 ml R2	99	36 months	cerebrospinal fluid	10.86 to 494.33 (mg/l)
	MATUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	18 months
------------------	-----------	--------	-----------

CONTROLS

H	MACOZ-002	1x2 ml	18 months
H	MACOX-002	1x2 ml	18 months
M	MACON-002	1x2 ml	18 months
L	MACOS-002	1x2 ml	18 months

ALPHA-1 ACID-GLYCOPROTEIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
KIT	GPTUR-B00/ATE	1 x 1 x 18,4 ml R1 + 1 x 1 x 5,7 ml R2	100	36 months	serum	0.143 to 2.42 (g/l)
	GPTUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7 ml R2	263			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
------------------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

ALPHA-1 ANTITRYPSIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	ATTUR-B00/ATE	1 x 1 x 13,2 ml R1 + 1 x 1 x 4,6 ml R2	99	36 months	serum or plasma	0.007 to 3.5 (g/l)
	ATTUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
------------------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-1 MICROGLOBULIN (URINE)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	A1DGX-B00/ATE	1 x 1 x 11 ml R1 + 1 x 1 x 11 ml R1	100	6 months	urine	1.29 to 166.6 (mg/l)
	A1DGX-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 24 ml R2	475			

CALIBRATORS

KIT 5 LEV	A1REK-000	5x1 ml	6 months
------------------	-----------	--------	----------

CONTROLS

M	A1CON-002	1x2 ml	6 months
L	A1COS-002	1x2 ml	6 months

ALPHA-2 MACROGLOBULIN SERUM and URINARY

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	MGTUR-B00/ATE	1 x 1 x 13,1 ml R1 + 1 x 1 x 4,6 ml R2	99	36 months	serum or plasma or urine	0.026 to 4.96 (g/l) serum 3.36 to 199 (mg/l) urine
	MGTUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7,5 ml R2	385			
	MGTUR-L00/ATE	5 x 2 x 24 ml R1 + 5 x 2 x 7,5 ml R2	1923			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
------------------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months
URINARY LOW	MGCU1-002	1x2 ml	12 months
URINARY HIGH	MGCU2-002	1x2 ml	12 months

CATALOGUE

ANTI-STREPTOLYSIN O

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ASLTX-B00/ATE	1 x 1 x 13,2 ml R1	1 x 1 x 3,9 ml R1	101	20 months	serum or plasma	19.99 to 864.85 (IU/ml)
	ASLTX-H00/ATE	1 x 2 x 24 ml R1	1 x 2 x 6,1 ml R2	387			

CALIBRATORS

KIT 5 LEV	ASREK-000	5x1 ml	12 months
-----------	-----------	--------	-----------

CONTROLS

M	ASCON-002	1x2 ml	12 months
---	-----------	--------	-----------

ANTITHROMBIN III

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	A3TUR-B00/ATE	1 x 1 x 13,1 ml R1	1 x 1 x 4,6 ml R1	99	36 months	plasma	28.15 to 806.30 (mg/l)
	A3TUR-H00/ATE	1 x 2 x 24 ml R1	1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

APOLIPOPROTEIN A1

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	AATUR-B00/ATE	1 x 1 x 13,2 ml R1	1 x 1 x 5,2 ml R1	99	36 months	serum or plasma	0.029 to 4.0 (g/l)
	AATUR-H00/ATE	1 x 2 x 24 ml R1	1 x 2 x 8,7 ml R2	383			

CALIBRATORS

KIT 5 LEV	ABREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	24 months
M	ABCON-002	1x2 ml	24 months

CATALOGUE

APOLIPOPROTEIN B

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	ABTUR-B00/ATE	1 x 1 x 13,2 ml R1 + 1 x 1 x 4,6 ml R1	99	36 months	serum or plasma	0.019 to 4.0 (g/l)
	ABTUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	ABREX-000	5 x 0,5 ml	18 months
------------------	-----------	------------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

BETA2-MICROGLOBULIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	B2DGX-B00/ATE	1 x 1 x 13 ml R1 + 1 x 1 x 6,7 ml R2	99	15 months	serum or plasma	0.132 to 95 (mg/l)
	B2DGX-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 12,7 ml R2	427			
	B2DGX-L00/ATE	5 x 2 x 24 ml R1 + 5 x 2 x 12,7 ml R2	2137			

CALIBRATORS

KIT 5 LEV	B2REK-000	5x1 ml	18 months
------------------	-----------	--------	-----------

CONTROLS

H	B2COX-002	1x2 ml	18 months
M	B2CON-002	1x2 ml	18 months
L	B2COS-002	1x2 ml	18 months

CERULOPLASMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	CETUR-B00/ATE	1 x 1 x 13,2 ml R1 + 1 x 1 x 4,6 ml R2	99	36 months	serum or plasma	9.48 to 1051.20 (mg/l)
	CETUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
------------------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

COMPLEMENT C3

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	C3TUR-B00/ATE	1 x 1 x 13,2 ml R1	1 x 1 x 4,6 ml R2	99	36 months	serum or plasma	0.015 to 4.31 (g/l)
	C3TUR-H00/ATE	1 x 2 x 24 ml R1	1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C4

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	C4TUR-B00/ATE	1 x 1 x 13,2 ml R1	1 x 1 x 4,6 ml R2	99	36 months	serum or plasma	0.002 to 0.825 (g/l)
	C4TUR-H00/ATE	1 x 2 x 24 ml R1	1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CRP GOLD LATEX

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CPDGX-B00/ATE	1 x 1 x 7 ml R1	1 x 1 x 7 ml R2	105	24 months	serum or plasma	0.031 to 300 (mg/l)
	CPDGX-H00/ATE	1 x 2 x 24 ml R1	1 x 2 x 24 ml R2	859			
	CPDGX-L00/ATE	5 x 2 x 24 ml R1	5 x 2 x 24 ml R2	4294			

CALIBRATORS

KIT 5 LEV	CPRLK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

CATALOGUE

CRP WIDE RANGE

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CPTXL-B00/ATE	1 x 1 x 12 ml R1 + 1 x 1 x 3,8 ml R2		97	36 months	serum or plasma	0.607 to 425 (mg/l)
	CPTXL-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 6,5 ml R2		421			
	CPTXL-L00/ATE	5 x 2 x 24 ml R1 + 5 x 2 x 6,5 ml R2		2103			

CALIBRATORS

KIT 5 LEV	CPREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	18 months
M	CPCON-002	1x2 ml	18 months
L	CXCON-002	1x2 ml	18 months

CYSTATIN C

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CYCOL-B00/ATE	1 x 1 x 10,1 ml R1 + 1 x 1 x 5,7 ml R2		100	6 months	serum or plasma	0.15 to 9.36 (mg/l)
	CYCOL-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 12,3 ml R2		517			

CALIBRATORS

KIT 5 LEV	CYREK-000	5x1 ml	12 months
-----------	-----------	--------	-----------

CONTROLS

H	CYCOX-002	1x2 ml	12 months
M	CYCON-002	1x2 ml	12 months
L	CYCOX-002	1x2 ml	12 months

FERRITIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	FNCOL-B00/ATE	1 x 1 x 10,5 ml R1 + 1 x 1 x 4,1 ml R2		100	12 months	serum	6.344 to 500 (µg/l)
	FNCOL-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 8 ml R2		502			

CALIBRATORS

KIT 5 LEV	FNRGK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

M	FNCON-002	1x2 ml	18 months
---	-----------	--------	-----------

CATALOGUE

FIBRINOGEN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	FITUR-B00/ATE	1 x 1 x 13,1 ml R1 + 1 x 1 x 5,7 ml R2	100	36 months	plasma (from sodium citrate)	0.05 to 10 (g/l)
	FITUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 10 ml R2	387			

CALIBRATORS

KIT	FIREM-001	1x1 ml	12 months
-----	-----------	--------	-----------

CONTROLS

M	FICON-002	1x2 ml	12 months
---	-----------	--------	-----------

HAPTOGLOBIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	HATUR-B00/ATE	1 x 1 x 13,1 ml R1 + 1 x 1 x 4,6 ml R2	99	36 months	serum or plasma	0.002 to 3.86 (g/l)
	HATUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS A

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	IATUR-B00/ATE	1 x 1 x 13,1 ml R1 + 1 x 1 x 4,6 ml R2	99	36 months	serum	0.02 to 8.0 (g/l)
	IATUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS E

NEW

Contact us

CATALOGUE

IMMUNOGLOBULINS G

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IGTUR-B00/ATE	1 x 1 x 13,1 ml R1	+ 1 x 1 x 6,7 ml R2	99	36 months	serum	0.125 to 35 (g/l)
	IGTUR-H00/ATE	1 x 2 x 24 ml R1	+ 1 x 2 x 7,6 ml R2	382			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS G (CSF)

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IGLCR-B00/ATE	1 x 1 x 11 ml R1	+ 1 x 1 x 4,1 ml R2	100	36 months	cerebrospinal fluid	3,375 to 249,78 (mg/l)
	IGLCR-H00/ATE	1 x 2 x 24 ml R1	+ 1 x 2 x 7,6 ml R2	471			

CALIBRATORS

KIT 5 LEV	IGREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	IGCOX-002	1x2 ml	24 months
M	IGCON-002	1x2 ml	24 months
L	IGCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS M

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IMTUR-B00/ATE	1 x 1 x 13,1 ml R1	+ 1 x 1 x 4,6 ml R2	99	36 months	serum	0.038 to 7.0 (g/l)
	IMTUR-H00/ATE	1 x 2 x 24 ml R1	+ 1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

LIPOPROTEIN A

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
		<i>R1</i>	<i>R2</i>			
KIT	LPTUR2-B00/ATE	1 x 1 x 13,1 ml R1 + 1 x 1 x 3,8 ml R2	97	12 months	serum or plasma	0.0235 to 0.946 (g/l)
	LPTUR2-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 6,1 ml R2	387			

CALIBRATORS

KIT 5 LEV	LPREK-000	5x1 ml	10 months
-----------	-----------	--------	-----------

CONTROLS

M	LPCON-002	1x2 ml	12 months
---	-----------	--------	-----------

MICROALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life*	Sample**	Analytical Range
		<i>R1</i>	<i>R2</i>			
KIT	MATUR-B00/ATE	1 x 1 x 13,1 ml R1 + 1 x 1 x 4,6 ml R2	99	36 months	urine	1.86 to 400 (mg/l)
	MATUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7,5 ml R2	385			

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	MACOX-002	1x2 ml	18 months
M	MACON-002	1x2 ml	18 months
L	MACOS-002	1x2 ml	18 months

PREALBUMIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		<i>R1</i>	<i>R2</i>			
KIT	PATUR-B00/ATE	1 x 1 x 18,5 ml R1 + 1 x 1 x 5,7 ml R2	100	36 months	serum or plasma	6.53 to 1000 (mg/l)
	PATUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 7 ml R2	263			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CATALOGUE

RETINOL BINDING PROTEIN (RBP)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	RBTUR-B00/ATE	1 x 1 x 11 ml R1 + 1 x 1 x 5,3 ml R2	98	24 months	serum or plasma	7,81 to 141 (mg/l)
	RBTUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 10,6 ml R2	471			

CALIBRATORS

KIT 5 LEV	RBREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	RBCOX-002	1x2 ml	24 months
M	RBCON-002	1x2 ml	24 months
L	RBCOS-002	1x2 ml	24 months

RHEUMATOID FACTOR

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	RFCOL-B00/ATE	1 x 1 x 10,5 ml R1 + 1 x 1 x 7,8 ml R2	100	18 months	serum or plasma	2.228 to 120 (kIU/l)
	RFCOL-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 17,3 ml R2	502			

CALIBRATORS

KIT 5 LEV	RFREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	RFCOX-002	1x2 ml	18 months
M	RFCON-002	1x2 ml	24 months
L	RFCOS-002	1x2 ml	18 months

TRANSFERRIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2			
KIT	TFTUR-B00/ATE	1 x 1 x 13,5 ml R1 + 1 x 1 x 5,7 ml R2	103	36 months	serum or plasma	0.027 to 8.19 (g/l)
	TFTUR-H00/ATE	1 x 2 x 24 ml R1 + 1 x 2 x 9,2 ml R2	388			
	TFTUR-L00/ATE	5 x 2 x 24 ml R1 + 5 x 2 x 9,2 ml R2	1940			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months



Your expert in specific proteins

(*) Tests : Amount of real tests. May vary from a analyzer to another one.

() Shelf Life :** DiAgam provides to distributors a minimum of **12 months** .(except for Alpha1-Microglobulin and cystatin C). The time period written in this document shows the maximum shelf life time after manufacturing.

(*) Sample :** DiAgam may provide some applications for other samples, such as CSF, Urine, Veterany samples etc.... Please contact your distributor or us directly at mail@diagam.com or dial **+32 68 55 14 82**.

(**)**
Analytical range : For general information purposes only. More details can be found in the package insert (instructions for use). Please contact your distributor or us directly at mail@diagam.com or dial **+32 68 55 14 82** in order to get the updated data.



Medical device for in vitro diagnosis.



DIAGAM SA, *Rue du Parc Industriel 40, 7822 Ghislenghien, BELGIUM*



The products fulfill the requirements of European directive 98/79 EC concerning the medical devices of in vitro diagnosis.

© DiAgam. 2021. All rights reserved. All product names, registered trademarks, company names in this document remain the property of their respective owners. For general information purposes only. More details can be found in the package insert (instructions for use) and labelling. Not for the US market except for research use only.