

ADVIA® LINE

New

SARS-CoV-2 SEROLOGICAL ASSAY

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
----------	-----------	---------	--------	--------------	-----------	----------------------

Please contact us for more information

SARS-CoV-2 SEROLOGICAL ASSAY based on PECIA Colloidal Gold



ALBUMIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ALTUR-B00/ADV	1 x 30 ml R1 + 1 x 8 ml R2		300	36 months	serum & plasma	0.815 to 124.94 (g/l)
	ALTUR-L00/ADV	2 x 70 ml R1 + 2 x 18 ml R2		1400			
	ALTUR-L00/ADV	5 x 70 ml R1 + 5 x 18 ml R2		3500			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALBUMIN CSF							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	MATUR-B00/ADV	1 x 46 ml R1 + 1 x 8 ml R2		400	36 months	cerebrospinal fluid	10.86 to 494.33 (mg/l)
	MATUR-L00/ADV	2 x 50 ml R1 + 1 x 15 ml R2		750			

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MACOZ-002	1x2 ml	24 months
H	MACOX-002	1x2 ml	24 months
M	MACON-002	1x2 ml	24 months
L	MACOS-002	1x2 ml	24 months

ALPHA-1 GLYCOPROTEIN ACIDE/OROSOMUCOID							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	GPTUR-B00/ADV	1 x 41 ml R1 + 1 x 8 ml R2		267	36 months	serum	0.143 to 2.42 (g/l)
	GPTUR-H00/ADV	2 x 70 ml R1 + 2 x 12 ml R2		800			
	GPTUR-C00/ADV	2 x 35 ml R1 + 2 x 8 ml R2		467			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-1 ANTITRYPSIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ATTUR-B00/ADV	1 x 46 ml R1 + 1 x 8 ml R2		400	36 months	serum & plasma	0.007 to 3.5 (g/l)
	ATTUR-H00/ADV	2 x 35 ml R1 + 2 x 9 ml R2		700			
	ATTUR-L00/ADV	1 x 70 ml R1 + 1 x 14 ml R2		700			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-1 MICROGLOBULIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	A1DGX-B00/ADV	1 x 15 ml R1 + 1 x 15 ml R2		375	17 months	urine	1.29 to 166.6 (mg/l)
	A1DGX-H00/ADV	2 x 20 ml R1 + 2 x 20 ml R2		1000			

CALIBRATORS

KIT 5 LEV	A1REK-000	5x1 ml	14 months
-----------	-----------	--------	-----------

CONTROLS

M	A1CON-002	1x2 ml	14 months
L	A1COS-002	1x2 ml	14 months

ALPHA-2 MACROGLOBULIN - ALPHA-2 MACROGLOBULIN URINARY

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	MGTUR-H00/ADV	1 x 50 ml R1 + 1 x 10 ml R2		500	36 months	serum & plasma urine	0.026 to 4.96 (g/l) 3.36 to 199 (mg/l)
	MGTUR-C00/ADV	2 x 35 ml R1 + 2 x 9 ml R2		700			
	MGTUR-L00/ADV	12 x 70 ml R1 + 12 x 12 ml R2		7200			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months
URINARY LOW	MGCU1-002	1x2 ml	24 months
URINARY HIGH	MGCU2-002	1x2 ml	24 months

ANTI STREPTOLYSIN O (ASO)

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ASLTX-B00/ADV	1 x 62 ml R1 + 1 x 8 ml R2		615	20 months	serum & plasma	19.99 to 864.85 (IU/ml)
	ASLTX-H00/ADV	2 x 70 ml R1 + 2 x 15 ml R2		1400			

CALIBRATORS

H	ASREK-000	5x1 ml	18 months
---	-----------	--------	-----------

CONTROLS

M	ASCON-002	1x2 ml	18 months
---	-----------	--------	-----------

APOLIPOPROTEIN A1

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	AATUR-B00/ADV	1 x 46 ml R1 + 1 x 12 ml R2		444	36 months	serum & plasma	0.029 to 4.0 (g/l)
	AATUR-H00/ADV	2 x 70 ml R1 + 2 x 20 ml R2		1400			

CALIBRATORS

KIT 5 LEV	ABREK-000	5x0.5 ml	18 months
-----------	-----------	----------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

APOLIPOPROTEIN B

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ABTUR-B00/ADV	1 x 46 ml R1 + 1 x 8 ml R2		400	36 months	serum & plasma	0.019 to 4.0 (g/l)
	ABTUR-H00/ADV	2 x 70 ml R1 + 2 x 18 ml R2		1400			

CALIBRATORS

KIT 5 LEV	ABREK-000	5x0.5 ml	18 months
-----------	-----------	----------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

BETA 2 MICROGLOBULIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	B2DGX-B00/ADV	1 x 20 ml R1 + 1 x 8 ml R2		200	30 months	serum & plasma	0.132 to 95 (mg/l)
	B2DGX-H00/ADV	1 x 44 ml R1 + 1 x 19 ml R2		475			
	B2DGX-L00/ADV	12 x 44 ml R1 + 12 x 19 ml R2		5700			

CALIBRATORS

KIT 5 LEV	B2REK-000	5x1 ml	30 months
-----------	-----------	--------	-----------

CONTROLS

H	B2COX-001	1 x 1 ml	30 months
M	B2CON-003	1x3 ml	30 months
L	B2COS-003	1x3 ml	30 months

BETA 2 MICROGLOBULIN (Urine)

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	BUDGX-B00/ADV BUDGX-H00/ADV BUDGX-L00/ADV	Contact us			30 months	Urine	0.055 to 9.3 (mg/l)

CALIBRATORS

KIT 5 LEV	BUREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

M	B2CON-003	1x3 ml	30 months
L	B2COS-003	1x3 ml	30 months

CALPROTECTIN FECAL

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CACOL-B00/ADV CACOL-H00/ADV	Contact us		101 303	18 months	Faecal samples	32,4 to 802,0 (mg/kg)

CALIBRATORS

KIT 6 LEV	CAREK-000	6x1 ml	12 months
-----------	-----------	--------	-----------

CONTROLS

H	CACOX-002	1x2 ml	12 months
M	CACON-002	1x2 ml	12 months
L	CACOS-002	1x2 ml	12 months

SAMPLE DILUTION BUFFER

KIT	SDBUF-B00	2x70 ml	15 months
	SDBUF-H00	8x70ml	15 months
	SDBUF-L00	16x70 ml	15 months

CALPRO EASY EXTRACT™

KIT	CAL0510	50 devices
-----	---------	------------

CERULOPLASMIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CETUR-B00/ADV	1 x 46 ml R1 + 1 x 8 ml R2		400	36 months	serum & plasma	9.48 to 1051.20 (mg/l)
	CETUR-H00/ADV	2 x 35 ml R1 + 2 x 9 ml R2		700			
	CETUR-L00/ADV	2 x 70 ml R1 + 2 x 18 ml R2		1400			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C3							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	C3TUR-B00/ADV	1 x 46 ml R1 + 1 x 8 ml R2		400	36 months	serum & plasma	0.015 to 4.31 (g/l)
	C3TUR-C00/ADV	2 x 35 ml R1 + 2 x 9 ml R2		700			
	C3TUR-H00/ADV	2 x 70 ml R1 + 2 x 18 ml R2		1400			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C4							
REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	C4TUR-B00/ADV	1 x 46 ml R1 + 1 x 8 ml R2		400	36 months	serum & plasma	0.002 to 0.825 (g/l)
	C4TUR-C00/ADV	2 x 35 ml R1 + 2 x 9 ml R2		700			
	C4TUR-H00/ADV	2 x 70 ml R1 + 2 x 18 ml R2		1400			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CRP GOLD LATEX

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CPDGX-B00/ADV	1 x 10 ml R1 + 1 x 10 ml R2		250	24 months	serum & plasma	0.031 to 300 (mg/l)
	CPDGX-H00/ADV	2 x 20 ml R1 + 2 x 20 ml R2		1000			
	CPDGX-L00/ADV	12 x 20 ml R1 + 12 x 20 ml R2		6000			

CALIBRATORS

KIT 5 LEV	CPRLK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	24 months
M	CPCON-002	1x2 ml	24 months
L	CXCON-002	1x2 ml	24 months

CRP XL WIDE RANGE

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CPTXL-B00/ADV	1 x 57 ml R1 + 1 x 8 ml R2		615	36 months	serum & plasma	0.607 to 425 (mg/l)
	CPTXL-H00/ADV	2 x 70 ml R1 + 2 x 10 ml R2		1538			
	CPTXL-L00/ADV	12 x 70 ml R1 + 12 x 15 ml R2		9333			

CALIBRATORS

KIT 5 LEV	CPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	24 months
M	CPCON-002	1x2 ml	24 months
L	CXCON-002	1x2 ml	24 months

CYSTATIN C - Colloidal Gold

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CYCOL-B00/ADV	1 x 25,5 ml R1 + 1 x 9 ml R2		300	24 months	serum & plasma	0.15 to 9.36 (mg/l)
	CYCOL-H00/ADV	1 x 50 ml R1 + 1 x 20,5 ml R2		680			

CALIBRATORS

KIT 5 LEV	CYREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	CYCOX-002	1x2 ml	24 months
M	CYCON-002	1x2 ml	24 months
L	CYCOS-002	1x2 ml	24 months

FERRITIN - Colloidal Gold

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	FNCOL-B00/ADV	1 x 40 ml R1 + 1 x 8 ml R2		533	12 months	serum	6.344 to 500 (µg/l)
	FNCOL-H00/ADV	2 x 70 ml R1 + 2 x 14 ml R2		1867			

CALIBRATORS

KIT 5 LEV	FNRGK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

M	FNCON-002	1x2 ml	18 months
---	-----------	--------	-----------

HAPTOGLOBIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	HATUR-B00/ADV	1 x 40 ml R1 + 1 x 8 ml R2		400	36 months	serum & plasma	0.002 to 3.86 (g/l)
	HATUR-C00/ADV	2 x 35 ml R1 + 2 x 9 ml R2		700			
	HATUR-H00/ADV	2 x 70 ml R1 + 2 x 18 ml R2		1400			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

HDL CHOLESTEROL

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	HDLC-KIT/ADV	12 x 60 ml R1 + 12 x 20 ml R2		Contact us	Contact us	serum	1 to 180 (mg/dl)

CALIBRATORS

KIT 1 LEV	BIOCAL	5x3 ml	Contact us
-----------	--------	--------	------------

CONTROLS

Control +	BIOPATH	6x5 ml	Contact us
Control -	BIONORM	6x5 ml	Contact us

IMMUNOGLOBULINS A

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	IATUR-B00/ADV	1 x 46 ml R1 + 1 x 8 ml R2		400	36 months	serum	0.02 to 8.0 (g/l)
	IATUR-H00/ADV	2 x 70 ml R1 + 2 x 18 ml R2		1400			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS E **NEW**

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	IEDGX-B00/ADV	Contact us			20 months	serum & plasma	21 to 1436 (U/ml)
	IEDGX-H00/ADV						

CALIBRATORS

KIT 5 LEV	IEREK-000	5x1 ml	20 months
-----------	-----------	--------	-----------

CONTROLS

M	IECON-002	1x2 ml	20 months
L	IECOS-002	1x2 ml	20 months

IMMUNOGLOBULINS G

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	IGTUR-B00/ADV	1 x 23 ml R1 + 1 x 8 ml R2		200	36 months	serum	0.125 to 35 (g/l)
	IGTUR-H00/ADV	2 x 50 ml R1 + 2 x 20 ml R2		1000			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS G (CSF)

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	IGLCR-B00/ADV	1 x 55 ml R1 + 1 x 10 ml R2		667	24 months	cerebrospinal fluid	3,375 to 249,78 (mg/l)
	IGLCR-H00/ADV	2 x 55 ml R1 + 2 x 10 ml R2		1333			

CALIBRATORS

KIT 5 LEV	IGREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	IGCOX-002	1x2 ml	24 months
M	IGCON-002	1x2 ml	24 months
L	IGCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS M

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	IMTUR-B00/ADV	1 x 46 ml R1 + 1 x 8 ml R2		400	36 months	serum	0.038 to 7.0 (g/l)
	IMTUR-H00/ADV	2 x 70 ml R1 + 2 x 18 ml R2		1400			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

LIPOPROTEIN (A)

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	LPTUR2-B00/ADV	1 x 70 ml R1 + 1 x 9,1 ml R2		700	12 months	serum & plasma	0.0235 to 0.946 (g/l)
	LPTUR2-H00/ADV	2 x 70 ml R1 + 2 x 9,1 ml R2		1400			

CALIBRATORS

KIT 4 LEV	LPREK-000	4x0,5 ml	18 months
-----------	-----------	----------	-----------

CONTROLS

M	LPCON-002	1x2 ml	24 months
---	-----------	--------	-----------

MICROALBUMIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	MATUR-B00/ADV	1 x 46 ml R1 + 1 x 8 ml R2		400	36 months	urine	1.86 to 400 (mg/l)
	MATUR-L00/ADV	2 x 50 ml R1 + 1 x 15 ml R2		750			

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MACOX-002	1x2 ml	24 months
M	MACON-002	1x2 ml	24 months
L	MACOS-002	1x2 ml	24 months

PREALBUMIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	PATUR-B00/ADV	1 x 41 ml R1 + 1 x 8 ml R2		267	36 months	serum & plasma	6.53 to 1000 (mg/l)
	PATUR-C00/ADV	2 x 35 ml R1 + 2 x 8 ml R2		467			
	PATUR-H00/ADV	2 x 70 ml R1 + 2 x 15 ml R2		933			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

RETINOL BINDING PROTEIN (RBP)							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	RBTUR-B00/ADV	1 x 31 ml R1 + 1 x 10 ml R2		400	36 months	serum & plasma	7,81 to 141 (mg/l)
	RBTUR-H00/ADV	2 x 31 ml R1 + 2 x 10 ml R2		800			
	RBTUR-L00/ADV	12 x 31 ml R1 + 12 x 10 ml R2		4800			

CALIBRATORS

KIT 5 LEV	RBREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	RBCOX-002	1x2 ml	18 months
M	RBCON-002	1x2 ml	18 months
L	RBCOS-002	1x2 ml	18 months

RHEUMATOID FACTOR (RF) - Colloidal Gold

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	RFCOL-B00/ADV	1 x 18 ml R1 + 1 x 10 ml R2		200	18 months	serum & plasma	2.228 to 120 (kIU/l)
	RFCOL-C00/ADV	2 x 40 ml R1 + 2 x 20 ml R2		800			
	RFCOL-H00/ADV	2 x 34,5 ml R1 + 2 x 20 ml R2		800			

CALIBRATORS

KIT 5 LEV	RFREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	RFCOX-002	1x2 ml	24 months
M	RFCON-002	1x2 ml	24 months
L	RFCOS-002	1x2 ml	24 months

TRANSFERRIN

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	TFTUR-C00/ADV	1 x 30 ml R1 + 1 x 8 ml R2		286	36 months	serum & plasma	0.027 to 8.19 (g/l)
	TFTUR-H00/ADV	2 x 70 ml R1 + 2 x 20 ml R2		1400			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months



(*) Tests : Amount of theoretical tests. May vary from a analyzer to another one.

() Shelf Life :** Diagam provides to distributors a minimum of *12 months* . The time period written in this document shows the maximum shelf life time after manufacturing.

(*) Sample :** Diagam may provide some applications for other samples, such as LCR, Urin, Veterinary samples etc.... Please contact your distributor or us directly at mail@diagam.com or dial *+32 68 55 14 82*.

(**)**
Analytical range : For general information purposes only. More details can be found in the package insert (instructions for use). Please contact your distributor or us directly at mail@diagam.com or dial *+32 68 55 14 82* in order to get the updated data.

 IVD Medical device for in vitro diagnosis.

 DIAGAM SA, *Rue du Parc Industriel 40, 7822 Ghislenghien, BELGIUM*

 The products fulfill the requirements of European directive 98/79 EC concerning the medical devices of in vitro diagnosis.

© Diagam. 2022. All rights reserved. All product names, registered trademarks, company names in this document remain the property of their respective owners. For general information purposes only. More details can be found in the package insert (instructions for use) and labelling. Not for the US market except for research use only.