

COBAS® LINE

New

SARS-CoV-2 SEROLOGICAL ASSAY

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
-----------------	------------------	----------------	---------------	---------------------	------------------	-----------------------------

Please contact us for more information

SARS-CoV-2 SEROLOGICAL ASSAY based on PECIA Colloidal Gold



ALBUMIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ALTUR-B00/COB	1 x 25,3 ml R1 + 1 x 18,2 ml R1 + 1 x 12,2 ml R2		230	36 months	serum & plasma	0.546 to 132.48 (g/l)
	ALTUR-H00/COB	2 x 25,3 ml R1 + 2 x 18,2 ml R1 + 2 x 12,2 ml R2		460			
	ALTUR-L00/COB	12 x 25,3 ml R1 + 12 x 18,2 ml R1 + 12 x 12,2 ml R2		2760			
	ALTUR-C00/COB	12 x 25,3 ml R1 + 12 x 18,2 ml R1 + 12 x 12,2 ml R2		2760			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALBUMIN CSF							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	MATUR-B00/COB	1 x 25,3 ml R1 + 1 x 18,2 ml R1 + 1 x 9,8 ml R2		230	36 months	CSF	10.86 to 494.33 (mg/l)
	MATUR-H00/COB	2 x 25,3 ml R1 + 2 x 18,2 ml R1 + 2 x 9,8 ml R2		460			
	MATUR-L00/COB	12 x 25,3 ml R1 + 12 x 18,2 ml R1 + 12 x 9,8 ml R2		2760			

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MACOZ-002	1x2 ml	24 months
H	MACOX-002	1x2 ml	24 months
M	MACON-002	1x2 ml	24 months
L	MACOS-002	1x2 ml	24 months

ALPHA-1 GLYCOPROTEIN ACIDE/OROSOMUCOID							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	GPTUR-B00/COB	1 x 22,2 ml R1 + 1 x 14,3 ml R1 + 1 x 8,3 ml R		150	36 months	serum	0.143 to 2.42 (g/l)
	GPTUR-H00/COB	2 x 22,2 ml R1 + 2 x 14,3 ml R1 + 2 x 8,3 ml R		300			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-1 ANTITRYPSIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	ATTUR-B00/COB	1 x 23,5ml R1 + 1 x 6,2 ml R2		100	36 months	serum & plasma	0.007 to 3.5 (g/l)
	ATTUR-H00/COB	2 x 23,5ml R1 + 2 x 6,2 ml R2		200			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

ALPHA-2 MACROGLOBULIN - ALPHA-2 MACROGLOBULIN URINARY

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	MGTUR-B00/COB	1 x 23,5ml R1 + 1 x 6,4 ml R2		100	36 months	serum & plasma or urine	0.026 to 4.96 (g/l)
	MGTUR-H00/COB	2 x 23,5 ml R1 + 2 x 6,2 ml R2		200			serum
	MGTUR-L00/COB	12 x 25,1 ml R1 + 12 x 17,2 ml R1 + 12 x 9,4 ml R2		2220			3.36 to 199 (mg/)
	MGTUR-C00/COB	12 x 23,5 ml R1 + 12 x 6,2 ml R2		1200			urine

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months
URINARY LOW	MGCU1-002	1x2 ml	24 months
URINARY HIGH	MGCU2-002	1x2 ml	24 months

ANTI STREPTOLYSIN O (ASO)

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****	
		R1	R2				
KIT	ASLTX-B00/COB	1 x 23,5 ml R1 + 1 x 16,9 ml R1 + 1 x 6,8 ml R2		180	20 months	serum & plasma	19.99 to 864.85 (IU/ml)
	ASLTX-H00/COB	2 x 23,5 ml R1 + 2 x 16,9 ml R1 + 2 x 6,8 ml R2		360			

CALIBRATORS

H	ASREK-000	5x1 ml	18 months
---	-----------	--------	-----------

CONTROLS

M	ASCON-002	1x2 ml	20 months
---	-----------	--------	-----------

APOLIPOPROTEIN A1

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	AATUR-B00/COB	1 x 22,2 ml R1 + 1 x 14,3 ml R1 + 1 x 10,4 ml R2		150	36 months	serum & plasma	0.029 to 4 (g/l)
	AATUR-H00/COB	2 x 22,2 ml R1 + 2 x 14,3 ml R1 + 2 x 10,4 ml R2		300			

CALIBRATORS

KIT 5 LEV	ABREK-000	5x0,5 ml	18 months
-----------	-----------	----------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

APOLIPOPROTEIN B

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	ABTUR-B00/COB	1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 9,1 ml R2		186	36 months	serum & plasma	0.019 to 4.0 (g/l)
	ABTUR-H00/COB	3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 9,1 ml R2		558			

CALIBRATORS

KIT 5 LEV	ABREK-000	5x0,5 ml	18 months
-----------	-----------	----------	-----------

CONTROLS

H	ABCOX-002	1x2 ml	18 months
M	ABCON-002	1x2 ml	18 months

BETA 2 MICROGLOBULIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	B2DGX-B00/COB	1 x 24,8 ml R1 + 1 x 16,4 ml R1 + 1 x 17,7 ml R2		240	30 months	serum & plasma	0,132 to 95 (mg/l)
	B2DGX-H00/COB	2 x 24,8 ml R1 + 2 x 16,4 ml R1 + 2 x 17,7 ml R2		480			
	B2DGX-L00/COB	12 x 24,8 ml R1 + 12 x 16,4 ml R1 + 12 x 17,7 ml R2		2880			

CALIBRATORS

KIT 5 LEV	B2REK-000	5x1 ml	30 months
-----------	-----------	--------	-----------

CONTROLS

H	B2COX-001	1 x 1 ml	30 months
M	B2CON-003	1x3 ml	30 months
L	B2COS-003	1x3 ml	30 months

BETA 2 MICROGLOBULIN (URINE)

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	BUDGX-B00/COB	1 x 13,9 ml R1 + 1 x 6,7 ml R2		100	30 months	Urine	0.055 to 9.3 (mg/l)
	BUDGX-H00/COB	2 x 13,9 ml R1 + 2 x 6,7 ml R2		200			

CALIBRATORS

KIT 5 LEV	BUREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

M	B2CON-003	1x3 ml	30 months
L	B2COS-003	1x3 ml	30 months

CALPROTECTIN FECAL

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CACOL-B00/COB	1 x 13,8 ml R1 + 1 x 6,4 ml R2		101	18 months	Faecal samples	32,4 to 802,0 (mg/kg)
	CACOL-H00/COB	3 x 13,8 ml R1 + 3 x 6,4 ml R2		303			

CALIBRATORS

KIT 6 LEV	CAREK-000	6x1 ml	12 months
-----------	-----------	--------	-----------

CONTROLS

H	CACOX-002	1x2 ml	12 months
M	CACON-002	1x2 ml	12 months
L	CACOS-002	1x2 ml	12 months

SAMPLE DILUTION BUFFER

KIT	SDBUF-B00	2x70 ml	15 months
	SDBUF-H00	8x70ml	15 months
	SDBUF-L00	16x70 ml	15 months

CALPRO EASY EXTRACT™

KIT	CAL0510	50 devices
-----	---------	------------

CERULOPLASMIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CETUR-B00/COB	1 x 23,5 ml R1 + 1 x 6,2 ml R2		120	36 months	serum & plasma	9.48 to 1051.20 (mg/l)
	CETUR-H00/COB	2 x 23,5 ml R1 + 2 x 6,2 ml R2		240			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C3

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	C3TUR-B00/COB	1 x 20,2 ml R1 + 1 x 13,0 ml R1 + 1 x 7,7 ml R2		133	36 months	serum & plasma	0.015 to 4.31 (g/l)
	C3TUR-H00/COB	3 x 20,2 ml R1 + 3 x 13,0 ml R1 + 3 x 7,7 ml R2		400			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

COMPLEMENT C4

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	C4TUR-B00/COB	1 x 23,5 ml R1 + 1 x 15,4 ml R1 + 1 x 8,7 ml R2		170	36 months	serum & plasma	0.002 to 0.825 (g/l)
	C4TUR-H00/COB	3 x 23,5 ml R1 + 3 x 15,4 ml R1 + 3 x 8,7 ml R2		510			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

CRP GOLD LATEX

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CPDGX-B00/COB	1 x 20,2 ml R1 + 1 x 18,3 ml R2		150	24 months	serum & plasma	0.257 to 370.56 (mg/l)
	CPDGX-H00/COB	3 x 20,2 ml R1 + 3 x 18,3 ml R2		450			
	CPDGX-L00/COB	12 x 20,2 ml R1 + 12 x 18,3 ml R2		1800			

CALIBRATORS

KIT 5 LEV	CPRLK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	24 months
M	CPCON-002	1x2 ml	24 months
L	CXCON-002	1x2 ml	24 months

CRP XL WIDE RANGE

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CPTXL-B00/COB	1 x 24,5 ml R1 + 1 x 15,8 ml R1 + 1 x 7,2 ml R2		189	36 months	serum & plasma	0.607 to 425 (mg/l)
	CPTXL-H00/COB	3 x 24,5 ml R1 + 3 x 15,8 ml R1 + 3 x 7,2 ml R2		566			
	CPTXL-L00/COB	12 x 24,5 ml R1 + 12 x 15,8 ml R1 + 12 x 7,2 ml R2		2264			

CALIBRATORS

KIT 5 LEV	CPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	CPCOX-002	1x2 ml	24 months
M	CPCON-002	1x2 ml	24 months
L	CXCON-002	1x2 ml	24 months

CYSTATIN C - Colloidal Gold

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	CYCOL-B00/COB	1 x 23,5 ml R1 + 1 x 10,5 ml R2		100	24 months	serum & plasma	0.15 to 9.355 (mg/l)
	CYCOL-H00/COB	2 x 23,5 ml R1 + 2 x 10,5 ml R2		200			
	CYCOL-L00/COB	12 x 23,5 ml R1 + 12 x 10,5 ml R2		1200			

CALIBRATORS

KIT 5 LEV	CYREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	CYCOX-002	1x2 ml	24 months
M	CYCON-002	1x2 ml	24 months
L	CYCOS-002	1x2 ml	24 months

FERRITIN - Colloidal Gold

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	FNCOL-B00/COB	1 x 21,6 ml R1 + 1 x 11,7 ml R1 + 1 x 7,7 ml R2		250	12 months	serum	6.344 to 500 (µg/l)
	FNCOL-H00/COB	2 x 21,6 ml R1 + 2 x 11,7 ml R1 + 2 x 7,7 ml R2		500			

CALIBRATORS

KIT 5 LEV	FNRGK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

M	FNCON-002	1x2 ml	18 months
---	-----------	--------	-----------

HAPTOGLOBIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	HATUR-B00/COB	1 x 22,2 ml R1 + 1 x 14,3 ml R1 + 1 x 8,3 ml R2		150	36 months	serum & plasma	0.002 to 3.86 (g/l)
	HATUR-H00/COB	2 x 22,2 ml R1 + 2 x 14,3 ml R1 + 2 x 8,3 ml R2		300			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS A

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IATUR-B00/COB	1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 9,1 ml R2		186	36 months	serum	0.02 to 8.0 (g/l)
	IATUR-H00/COB	3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 9,1 ml R2		558			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS G

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IGTUR-B00/COB	1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 14,9 ml R2		187	36 months	serum	0.125 to 35 (g/l)
	IGTUR-H00/COB	3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 14,9 ml R2		560			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS G (CSF)

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IGLCR-B00/COB	1 x 23,5 ml R1 + 1 x 12 ml R1 + 1 x 7,7 ml R2		200	36 months	CSF	3,375 to 249,78 (mg/l)
	IGLCR-H00/COB	2 x 23,5 ml R1 + 2 x 12 ml R1 + 2 x 7,7 ml R2		400			

CALIBRATORS

KIT 5 LEV	IGREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	IGCOX-002	1x2 ml	24 months
M	IGCON-002	1x2 ml	24 months
L	IGCOS-002	1x2 ml	24 months

IMMUNOGLOBULINS M

REAGENTS	Reference	Content		Tests*	Shelf Life*	Sample**	Analytical Range
		R1	R2				
KIT	IMTUR-B00/COB	1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 9,1 ml R2		186	36 months	serum	0.038 to 7.0 (g/l)
	IMTUR-H00/COB	3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 9,1 ml R2		558			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

LIPOPROTEIN (A)

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	LPTUR2-B00/COB	1 x 23,5 ml R1 + 1 x 17,2 ml R1 + 1 x 6,7 ml R2		200	12 months	serum or plasma	0,0235 to 0,9460 (g/l)
	LPTUR2-H00/COB	2 x 23,5 ml R1 + 2 x 17,2 ml R1 + 2 x 6,7 ml R2		400			

CALIBRATORS

KIT 4 LEV	LPREK-000	4x0,5 ml	18 months
-----------	-----------	----------	-----------

CONTROLS

M	LPCON-002	1x2 ml	24 months
---	-----------	--------	-----------

MICROALBUMIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	MATUR-B00/COB	1 x 25,2 ml R1 + 1 x 6,5 ml R2 + 1 x 4,9 ml R3		129	36 months	urine	4,30 to 476,93 (mg/l)
	MATUR-H00/COB	2 x 25,2 ml R1 + 2 x 6,5 ml R2 + 2 x 4,9 ml R3		258			
	MATUR-L00/COB	12 x 25,2 ml R1 + 12 x 6,5 ml R2 + 12 x 4,9 ml R2		1547			

CALIBRATORS

KIT 5 LEV	MAREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MACOX-002	1x2 ml	24 months
M	MACON-002	1x2 ml	24 months
L	MACOS-002	1x2 ml	24 months

PREALBUMIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	PATUR-B00/COB	1 x 24,3 ml R1 + 1 x 15,7 ml R1 + 1 x 9,1 ml R2		186	36 months	serum & plasma	6.53 to 1000 (mg/l)
	PATUR-H00/COB	3 x 24,3 ml R1 + 3 x 15,7 ml R1 + 3 x 9,1 ml R2		558			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

RETINOL BINDING PROTEIN							
REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	RBTUR-B00/COB	1 x 20,7 ml R1 + 1 x 7,7 ml R2		200	36 months	serum & plasma	7,81 to 141 (mg/l)
	RBTUR-H00/COB	2 x 20,7 ml R1 + 2 x 7,7 ml R2		400			
	RBTUR-L00/COB	12 x 20,7 ml R1 + 12 x 7,7 ml R2		2400			

CALIBRATORS

KIT 5 LEV	RBREK-000	5x1 ml	18 months
-----------	-----------	--------	-----------

CONTROLS

H	RBCOX-002	1x2 ml	18 months
M	RBCON-002	1x2 ml	18 months
L	RBCOS-002	1x2 ml	18 months

RHEUMATOID FACTOR (RF) - Colloidal Gold

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	RFCOL-B00/COB	1 x 18,2 ml R1 + 1 x 9,9 ml R1 + 1 x 16,7 ml R2		200	18 months	serum & plasma	2.228 to 120 (kUI/l)
	RFCOL-H00/COB	3 x 18,2 ml R1 + 3 x 9,9 ml R1 + 3 x 16,7 ml R2		600			

CALIBRATORS

KIT 5 LEV	RFREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	RFCOX-002	1x2 ml	24 months
M	RFCON-002	1x2 ml	24 months
L	RFCOS-002	1x2 ml	24 months

TRANSFERRIN

REAGENTS	Reference	Content		Tests*	Shelf Life**	Sample***	Analytical Range****
		R1	R2				
KIT	TFTUR-B00/COB	1 x 24,5 ml R1 + 1 x 16,7 ml R1 + 1 x 12,5 ml R2		180	36 months	serum & plasma	0.076 to 8.0 (g/l)
	TFTUR-H00/COB	1 x 24,5 ml R1 + 1 x 16,7 ml R1 + 1 x 12,5 ml R2		500			
	TFTUR-L00/COB	1 x 24,5 ml R1 + 1 x 16,7 ml R1 + 1 x 12,5 ml R2		2160			

CALIBRATORS

KIT 5 LEV	MPREK-000	5x1 ml	24 months
-----------	-----------	--------	-----------

CONTROLS

H	MPCOX-002	1x2 ml	24 months
M	MPCON-002	1x2 ml	24 months
L	MPCOS-002	1x2 ml	24 months

Please contact us for more information on the following tests:

ALPHA-1-MICROGLOBULIN

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
----------	-----------	---------	--------	--------------	-----------	----------------------

IMMUNOGLOBULIN E

REAGENTS	Reference	Content	Tests*	Shelf Life**	Sample***	Analytical Range****
----------	-----------	---------	--------	--------------	-----------	----------------------

Your expert in specific proteins



(*) Tests : Amount of real tests. May vary from a analyzer to another one.

() Shelf Life :** Diagam provides to distributors a minimum of *12 months*. The time period written in this document shows the maximum shelf life time after manufacturing.

(*) Sample :** Diagam may provide some applications for other samples, such as CSF, Urin, Veterany samples etc.... Please contact your distributor or us directly at mail@diagam.com or dial *+32 68 55 14 82*.

(**)**
Analytical range : For general information purposes only. More details can be found in the package insert (instructions for use). Please contact your distributor or us directly at mail@diagam.com or dial *+32 68 55 14 82* in order to get the updated data.



Medical device for in vitro diagnosis.



DIAGAM SA, *Rue du Parc Industriel 40, 7822 Ghislenghien, BELGIUM*



The products fulfill the requirements of European directive 98/79 EC concerning the medical devices of in vitro diagnosis.